

AirSim Pierre Robin X User Manual



An innovative and durable solution for realistic difficult infant airway management training

The AirSim® Pierre Robin X model (based on a 6-month old) was developed in collaboration with clinicians from Children's Hospital of Los Angeles.

This realistic training manikin highlights difficult airway management techniques for infants with Pierre Robin Sequence/Syndrome (PRS).

The nasal cavity and airway include important internal landmarks for training in endo-tracheal and nasotracheal intubations.

TruCorp are proud to offer a 5-year warranty on the AirSim X airway.

Contents

Product specifications	2
Package contents	2
Recommended equipment sizes	2
Initial set-up information	2
List of procedures facilitated by the AirSim Pierre Robin X	3
Care and maintenance	3
Warranty	4

Product specifications

Product Code: PR10000X

AirSim Pierre Robin X weight: Approx. 2kg

Shipment dimensions: 32 x 19 x 19cm

Package contents

- 1 AirSim Pierre Robin X ready-to-use (PR10000X)
- 1 AirSim Pierre Robin X carrier case (JBAG03)
- 100ml bottle of lubrication (TL001)
- 2 sets of adhesive pads (attachable to suction cups for increased stability if required)

Recommended equipment sizes

- 3.5-4.0 mm ID for oral intubation
- Size 1 for supraglottic devices
- Size 1 for video-laryngoscopy

Initial set-up information

- Gently remove the model from the carrier case
- Secure the manikin to a suitable surface by adjusting the screw on the suction cups (surface must be non-porous to ensure sufficient suction)
- Optional!: Adhesive pads are provided with the model for increased stability if required. This is particularly useful when the product is not on an ideal surface for suction cups to be effective. Simply peel the cover from the pad to locate the adhesive area and attach to the suction cups.
- The head can be placed in a neutral or sniffing position by rotating the head, if required for direct laryngoscopy

- Please ensure two lung bags and one stomach bag are attached to the airway and oesophagus respectively
- Spray generous amounts of lubrication provided with the model (TL001) on the airway and equipment prior to use

List of procedures facilitated by the AirSim Pierre Robin X

- Double naso-tracheal intubation
- Bag valve mask (BVM) ventilation techniques
- Full range of supraglottic devices
- Direct and video laryngoscopy
- Endotracheal tube insertion

Care and maintenance

Store in clean, dry conditions away from heat and direct sunlight; avoid contact with metals, solvents, oils or greases and strong detergents. When the product is not in use please store in the black carrier case provided.

Thoroughly wash the AirSim X airway in warm water. Please use warm soapy water or something similar until all visible foreign matter and residue is removed.

Mild detergents or enzymatic cleaning agents may be used on the airway in accordance with the manufacturer's instructions and at the proper dilution. The detergent must not contain skin or mucous membrane irritants.

Please do not use any of the following when cleaning the AirSim product range

- Germicides, disinfectants, or chemical agents such as glutaraldehyde (e.g. Cidex®),
- Ethylene oxide, phenol-based cleaners or iodine-containing cleaners

In response to the recent COVID-19 pandemic, we recommend this additional step to ensure the product is fully sanitised:

Use alcohol spray (minimum 75%) and wipe off. Repeat this for 3-4 times to ensure to kill the virus completely. This method can be used on both the silicone skin and the latex airway.

Warranty

TruCorp warrants this unit to be free of defects in materials and workmanship and to give satisfactory service for a period of 1-year from the date of delivery. This ensures that our customers receive maximum coverage on each product. If the unit should malfunction it must be returned to the factory for evaluation. Upon examination by TruCorp, if the unit is found to be defective it will be repaired or replaced at no charge.

Additionally, TruCorp warrants a 5-year warranty on TruCorp branded airways on any AirSim X, Trauma X and TruBaby X range. The 5-year warranty only covers the actual TruCorp airway and not any other part of the model.

TruCorp will pay for the freight/delivery and the actual parts needed free of charge if any part of the product fails within the 1-year period. TruCorp will pay for the freight/delivery of the TruCorp airway free of charge if the airway fails within the 5-year period.

However, these warranties are VOID, if; the unit shows evidence of having been tampered with or shows evidence of having been damaged by excessive heat, the use of sharp instruments, misapplication, misuse or other operating conditions outside of TruCorp's control. Components which wear or are damaged by misuse are not warranted and will be charged for if repair has been approved. Warranty is void if third party products are seen to have damaged or caused failure of the TruCorp models.

Please direct all warranty and repair inquiries to:



TruCorp Ltd
33 Waringstown Road
Lurgan, Co. Armagh
BT66 7HH
Northern Ireland
info@trucorp.com
Tel: +44 (0) 28 3888 2714